DATA SHEET

BCKP05 TYPE 'K' NEEDLE PROBE

Needle Probes - 3.3 mm Type 'K'

Description

This probe uses the bulbous handle to enable the sensor tip to be pushed into a semi-solid product with maximum ease of use.

Construction

Needle Probe 3.3mm Diameter by 115mm Long: Stainless Steel 316 (Food Grade) 1M curly polyurethane cable with moulded connector. Complete waterproof assembly.

Sensor Features

TOTAL ENCAPSULATION TECHNIQUE FOR MAXIMUM STRENGTH AND DURABILITY.

This results in a solid handle as opposed to a hollow handle. This is particularly important as there is often damage to the handles caused by excess heat. With a hollow handle it is possible to puncture the outer plastic and damage the sensor irreparably.

WATERPROOF HANDLE

Due to the total encapsulation method used, all TME probe handles are completely waterproof.

TOUGH POLYURETHANE CABLE

- Polyurethane cables are used in place of the standard PVC for the following reasons:-
- Non-Toxic
- · Greater mechanical strength for durability
- 12 X 0.2mm wires used internally for greater strength.
- PTFE inner insulation for strength and retractability.

HIGH ACCURACY THERMOCOUPLE MATERIAL THROUGHOUT

Type 'K' Thermocouple : Class II (±3.0°C ±0.5%)

POLYPROPYLENE HANDLES

Polypropylene is an extremely tough and durable material, commonly used for milk crates, it has good low temperature performance and a relatively high melt temperature. It performs exceptionally well under chemical attack.

➤ WIDE AMBIENT TEMPERATURE SPECIFICATION : -30 TO 50 °C
➤ TIME RESPONSE (96% of value in water) : 1.6 Secs
➤ MEASUREMENT RANGE : -100 TO 250 °C

Cross-reference for compatible instruments

Suitable instruments for use with this probe

TME PART No	DESCRIPTION	APPLICATION
MM2000	SINGLE INPUT INSTRUMENT	HIGH ACCURACY TEMPERATURE MEASUREMENT
MM2010	MAX / MIN HOLD INSTRUMENT	HIGH ACCURACY INSTRUMENT WITH MAX, MIN AND HOLD FEATURES
MM2020	DIFFERENTIAL INSTRUMENT	DUAL INPUT INSTRUMENT FOR DIFFERENTIAL MEASUREMENTS
MM2030	THERMOCOUPLE SIMULATOR	HIGH ACCURACY SIMULATOR WITH MEASUREMENT FACILITY